
Marine Physical Laboratory

Service Support for the Phillips Laboratory Whole Sky Imager

Janet E. Shields, Monette Karr and
Richard W. Johnson

Supported by the
Chief of Naval Research
Contract N00014-96-D-0065-DO#3

Final Report

MPL-U-10/97
July 1997

19971230 082

Approved for public release; distribution is unlimited.



University of California, San Diego
Scripps Institution of Oceanography

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.				
1. Agency Use Only (Leave Blank).		2. Report Date. July 1997		3. Report Type and Dates Covered. Final Report
4. Title and Subtitle. Service Support for the Phillips Laboratory Whole Sky Imager			5. Funding Numbers. N00014-96-D-0065-DO#3	
6. Author(s). Janet E. Shields, Monette Karr and Richard W. Johnson			Project No. Task No.	
7. Performing Monitoring Agency Name(s) and Address(es). University of California, San Diego Marine Physical Laboratory Scripps Institution of Oceanography San Diego, California 92152			8. Performing Organization Report Number. MPL-U-10/97	
9. Sponsoring/Monitoring Agency Name(s) and Address(es). Air Force Phillips Laboratory PL1610 3550 Aberdeen Drive SE, Bldg. 405 Kirtland AFB, NM 87117-5776 Atten: Capt. Jon Eric Claunch			10. Sponsoring/Monitoring Agency Report Number.	
11. Supplementary Notes.				
12a. Distribution/Availability Statement. Approved for public release; distribution is unlimited.			12b. Distribution Code.	
13. Abstract (Maximum 200 words). The Whole Sky Imager (WSI) is a ground-based digital imaging system for assessment of cloud cover over the full upper hemisphere. Using a fisheye lens and a slow scan CCD sensor, it acquires imagery under daylight, moonlight and starlight conditions. This contract funding enabled Marine Physical Laboratory to provide additional software development and documentation support for a new Whole Sky Imager which was delivered to the Air Force in December 95. The primary enhancements included several additions to the networking options, and delivery of a new "Operations Manual" and "Theory of Operations Manual".				
14. Subject Terms. digital imaging system, Whole Sky Imager, cloud cover, cloud, radiance, atmospheric			15. Number of Pages. 5	
			16. Price Code.	
17. Security Classification of Report. Unclassified	18. Security Classification of This Page. Unclassified	19. Security Classification of Abstract. Unclassified	20. Limitation of Abstract. None	

Service Support for the Phillips Laboratory Whole Sky Imager

**Janet E. Shields, Monette Karr and
Richard W. Johnson**

**Supported by the
Chief of Naval Research
Contract N00014-96-D-0065-DO#3
for the Period 4-22-96 - 4-28-97**

Abstract

The Whole Sky Imager (WSI) is a ground-based digital imaging system for assessment of cloud cover over the full upper hemisphere. Using a fisheye lens and a slow scan CCD sensor, it acquires imagery under daylight, moonlight and starlight conditions. This contract funding enabled Marine Physical Laboratory to provide additional software development and documentation support for a new Whole Sky Imager which was delivered to the Air Force in December 95. The primary enhancements included several additions to the networking options, and delivery of a new "Operations Manual" and "Theory of Operations".

Research Summary

The Day/Night Whole Sky Imager is an automated digital imaging system for acquiring imagery of the full upper hemisphere during both day and night. The WSI was developed by the Marine Physical Laboratory, based on many years experience with earlier daytime-only WSIs, which were also developed at MPL. At the time of the initiation of this contract, MPL had developed and fielded five Day/Night WSI units under funding from Air Force, DOE, Army, and Navy. The most recently delivered instrument was delivered to the Air Force's Philips Laboratory, under Contract N00014-93-D-0141-D11. The current contract enabled

MPL to develop additional networking capabilities, and deliver additional documentation.

Research Objectives

The contract was designed to allow MPL to complete an "Operations Manual" and "Theory of Operations Manual", and to enhance the features of the networking portions of the WSI archival software. Some of the documentation was originally funded in an earlier contract, but at the request of the sponsor, the work on this documentation was deferred to the present contract, and in return MPL provided software development well beyond the scope of the original contract. The enhancements to the networking software will be discussed below.

Discussion of the Results

This section discusses the new software and documentation which were delivered under this contract.

Software Development

At the beginning of this contract, the networking software was capable of transferring images to a remote display computer, transferring status information to the display computer, and receiving instructions from the display computer. The instructions allowed the user at the remote display computer to select which images were transferred, choose the archival interval (such as 10 min or 2 min between image sets), turn the tape archival feature on and off, and turn the processing options such as image ratio and cloud decision image on and off.

The Statement of Work was intentionally left flexible, so that MPL could meet changing needs as they developed, up to the limit of the available funding. As a result, the changes were made in several steps; the first requests were completed, and the program delivered, and then additional requests were completed as they developed. The primary changes are described below.

The flux control algorithm on the WSI normally determines the choice of filters and camera exposure, depending on ambient lighting conditions. This algorithm takes into account the position of the sun, and the

position, phase, and relative brightness of the moon. It is fast enough to choose appropriate settings even during sunrise and sunset when brightness levels are changing rapidly. While this algorithm is successful in choosing reasonable exposures, there are conditions in which the user might wish to change the exposure. For example, if monitoring a quickly changing event, the user might prefer shorter exposure times, which will yield poorer signal-to-noise statistics but better time response. The WSI was changed so that a user at the instrument can multiply the exposure time chosen by the algorithm by factors such as 10 or 0.1. By using a factor in this way, the algorithm still chooses a new exposure as the lighting conditions change (especially at sunrise and sunset), but the user can adjust the system to optimize image quality or time response.

The 16-bit raw data are normally displayed in 8-bit resolution on the monitor. A windowing algorithm chooses a reasonable range within the original 65,536 levels to map into the 256 levels inherent in the 8-bit image. The range selected by the algorithm is now written into the 16-bit image header, so that when the 16-bit image is transferred to the display system, the display system may use the same display range.

Two options were added for faster response. The program had previously allowed only two minute intervals as the minimum. This was changed to a one minute minimum. In addition, a fast grab feature was added, which allows the WSI to go as fast as it can (depending on the selected features), without waiting for an integer minute interval. The actual speed can be increased by turning off features such as archival to tape and cloud image processing, and it can be further enhanced by decreasing the exposure times. An additional feature was added, in which the WSI can be set to save every nth image set to the tape archive. This is used when the user wishes to save sample data, but also wants the fast response time.

Several options for occulter performance were added. The occulter can be set to disable the trolley drive functions; this could be used if the trolley is mechanically disabled. Another option allows the user to stow the occulter and not update the position of either the arc or the trolley. This has been used under starlight to block lights - the occulter is moved to the desired position manually, and then set to stay in place. A third change is that the occulter now goes to the east when there is no sun or moon.

The program has the option to use either the WWV clock or a GPS clock. An additional change was to allow the above options to be changed not just at the WSI, but also via the Scramnet from the display computer.

Conclusion

Making these software changes was somewhat challenging, because the system was no longer available to the personnel at MPL. Similar systems were available part of the time, but at no time was a system with Scramnet available. As a result, the programs had to be delivered with less than optimal testing. In spite of this limitation, the upgrades were made successfully with very few corrections required.

Documentation Enhancements

The new documentation delivered under this contract includes a new "Operations Manual", a "Theory of Operations", and documentation of the new software. The Operations Manual, Tech Note 240, includes new descriptions of the various subsystems and how they are designed to work, discussions of instrument operation, and setup instructions. This tech note was partially funded by another sponsor. The "Theory of Operations and Interactive Optimization", Tech Note 243, discusses the general theory of operations, the program sequence of events and timing, the flux control algorithm, the input file and hot key options, the cloud algorithm and derivation of its inputs, and the calibration theory. These Tech Notes are fairly substantial, with approximately 100 pages of documentation in each Tech Note.

Conclusion

The Whole Sky Imager software has been further developed to enable more interactive control of the instrument, both at the instrument controller and from a remote computer. The documentation has been upgraded to provide the user with a better understanding of the instrument and its use. Our understanding is that the sponsor is very pleased with both of these enhancements.

References

- Shields, J. E., R. W. Johnson, M. E. Karr, D. R. Sauer, and R. A. Weymouth (1996), "Operations Manual; Day/Night Whole Sky Imager", Technical Note No. 240, Atmospheric Optics Group, Marine Physical laboratory, University of California, San Diego
- Shields, J. E., R. W. Johnson, M. E. Karr, D. R. Sauer, J. R. Varah, and R. A. Weymouth (1996), "Maintenance and Trouble Shooting Manual; Day/

Acknowledgements

Night Whole Sky Imager", Technical Note No. 241, Atmospheric Optics Group, Marine Physical laboratory, University of California, San Diego

Shields, J. E., R. W. Johnson, M. E. Karr, D. R. Sauer, and R. A. Weymouth (1996), "Theory of Operations and Interactive Optimization; Day/Night Whole Sky Imager", Technical Note No. 243, Atmospheric Optics Group, Marine Physical laboratory, University of California, San Diego

Shields, J. E., R. W. Johnson, M. E. Karr, R. A. Weymouth, and D. S. Sauer (1997), "Development and Delivery of a Day/Night Whole Sky Imager with Enhanced Angular Alignment for Full 24 hour Cloud Distribution Assessment", Marine Physical laboratory, University of California, San Diego, report number to be determined

Acknowledgements

We would like to express our thanks to Capt. J. Eric Claunch of Philips Lab for his support and advice, to Gerry Williams for his input on the software development, and to ONR for their administration of this contract. Janet Shields

ONR/MPL REPORT DISTRIBUTION

Office of Naval Research (3)
Department of the Navy
Ballston Tower One
800 North Quincy Street
Arlington, VA 22217-5660
Attn: Code 321B

Air Force Phillips Laboratory
PL1610
3550 Aberdeen Drive SE, Bldg. 405
Kirtland AFB, NM 87117-5776

Regional Director (1)
ONR Detachment
San Diego Regional Office
4520 Executive Drive, Suite 300
San Diego, CA 92121-3019

Commanding Officer (1)
Naval Research Laboratory
4555 Overlook Avenue, S.W.
Attn: Code 2627
Washington, D.C. 20375-5320

Defense Technical Information Center (4)
8725 John J. Kingman Road
Suite 0944
Ft Belvoir, VA 22060-6218